



校准证书

CALIBRATION CERTIFICATE

证书编号 CYQ202211658
Certificate No.

第 1 页, 共 3 页
Page of

委托方 优利德科技(中国)股份有限公司
Client

委托方联络信息 广东省东莞市松山湖园区工业北一路6号
Contact Information

计量器具名称 望远镜测距仪
Description

型号/规格 LM1500G
Model/Type

制造厂 UNI-T
Manufacturer

出厂编号 自编22042004 设备管理编号 ----
Serial No. Equipment No.

接收日期 2022 年 04 月 01 日
Date of Receipt Y M D

结果 见校准结果
Results Shown in the results of calibration

校准日期 2022 年 04 月 20 日
Date of Calibration Y M D


批准人 陈伟琪
Approved Signatory

核 验 范斌斌
Reviewed by

校 准 孙彦锋
Calibrated by



扫一扫查真伪



说明

证书编号 CYQ202211658
Certificate No.

DIRECTIONS

第 2 页, 共 3 页
Page of

1. 本中心是国家市场监督管理总局在华南地区设立的国家法定计量检定机构, 本中心的质量管理体系符合 ISO/IEC 17025:2017 标准的要求。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the State Administration for Market Regulation. The quality system is in accordance with ISO/IEC 17025:2017.

2. 本中心所出具的数据均可溯源至国家计量基准和/或国际单位制(SI)。

All data issued by this laboratory are traceable to national primary standards and/or International System of Units (SI).

3. 校准地点、环境条件:

Place and environmental conditions of the calibration:

地点 本中心测绘仪器实验室

Place (Survey Instrument Lab.)

温度 (20.0 ± 3.0) °C

Temperature

相对湿度

(50 ± 5) %

R.H.

4. 本次校准的技术依据:

Reference documents for the calibration:

JJF1704-2018 望远镜式测距仪校准规范 C. S. for Telescope Rangefinders

5. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号规格

Name of Equipment

/Model/Type

标准钢卷尺

Standard Steel Tape

/100 m

编号

Serial No.

190006

证书号/有效期/溯源单位

Certificate No./Due Date

/Traceability to

CJC202110312

/2022-06-29

/本中心

计量特性

Metrological

Characteristic

$U = 5\mu\text{m} + 5 \times 10^{-6} L, k = 2$

白云机场标准长度基线场

Baiyun airport standard

length baseline field

/(24~1008) m

JX01等

JX2021-25号

/2024-08-24

/自然资源第一测量队

$U = (0.14 \sim 0.94) \text{mm} (k = 2)$

全站仪

Electronic Tachometer Total

Station

/TC2003

439658

CYQ202210532

/2023-02-14

/本中心

I 级

Class I

注: 1. 本证书校准结果只与受校准仪器有关。The results relate only to the items calibrated.

Note: 2. 未经本机构书面批准, 不得部分复制此证书。This certificate shall not be reproduced except in full, without the written approval of our laboratory.

3. “委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“出厂编号”以及“设备编号”为仪器上标注, 委托方对上面内容如有异议, 须在收到证书后二十个工作日内提出。

The information Client and Contact Information are provided by client, and the Manufacturer, Model/Type, Serial No. and Equipment No. are marked on the items. Client shall submit any objection within 20 working days after receiving the certificate for the information above.

4. 本次校准日期视为发布日期。The calibration date is the date of issue of the certificate.



校准结果 RESULTS OF CALIBRATION

证书编号 CYQ202211658
Certificate No.

原始记录号 CYQ202211658
Record No.

第 3 页, 共 3 页
Page of

- 1 外观质量及各项功能: 符合要求
Appearance and functions: Pass
- 2 测量重复性(标准偏差): 0.1 m
Repeatability of measurement (experimental standard deviation):
- 3 示值误差(长度):
Indication error(length):

受测点 (m) Point	示值误差(m) Indication error	受测点 (m) Point	示值误差(m) Indication error
10	-0.2	240	-2.2
20	-0.4	360	+0.6
30	-0.8	552	-1.1
72	-0.4	792	+0.7
144	0.0	1489	+3.0

说明:
Note:

- 1 本次室外检定地点: 白云机场标准长度基线场
Field place of the verification:

环境条件: 温度: 16℃
Environmental condition Temperature

- 2 示值误差(长度)测量结果的扩展不确定度: $U = 0.4 \text{ m}$ 包含因子 $k = 2$
Expanded uncertainty of measurement for indication error(length) Coverage factor

本证书中给出的扩展不确定度依据JJF1059.1-2012《测量不确定度评定与表示》评定, 由合成标准不确定度乘以包含概率约为95%时对应的包含因子 k 得到。

The expanded uncertainty given in this certificate is evaluated according to JJF 1059.1-2012 *Evaluation and Expression of Uncertainty in Measurement*, which is obtained by multiplying the combined standard uncertainty by the coverage factor k corresponding to the coverage probability of about 95%.

- 3 由于复校时间间隔的长短由仪器使用情况、使用者、仪器本身质量等诸因素所决定的, 因此, 送校单位可根据实际情况自主决定复校时间间隔。建议不超过1年。更换重要部件、维修或对仪器性能有怀疑时, 应及时校准。

Since the calibration interval is depended on a number of factors, such as the use of the instrument, operation of the user, and the quality of the instrument itself, the next calibration date can be decided by the user according to the actual use. Next calibration for this instrument is proposed within 1 year. When replacing important parts, repairs, or doubts about the performance of the instrument, it should be calibrated in time.