



校准证书

CALIBRATION CERTIFICATE

证书编号 GRRF202300043
Certificate No.

第 1 页, 共 3 页
Page of

委托方 优利德科技(中国)股份有限公司
Client

委托方联络信息 广东省东莞市松山湖园区工业北一路6号
Contact Information

计量器具名称 红外探针二合一测温仪(辐射温度计)
Description

型号规格 A63
Model/Type

制造厂 UNI-T
Manufacturer

出厂编号 C222759187
Serial No.

设备编号 /
Equipment No.

接收日期 2023 年 03 月 21 日
Date of Receipt Y M D

结果 见校准结果
Results

校准日期 2023 年 03 月 29 日
Date of calibration Y M D

批准人 胡启亮
Approved Signatory

核 验 蔡锡松 蔡锡松
Reviewed by

校 准 李礼爱 李礼爱
Calibrated by

证书专用章
Stamp





说明

证书编号 GRRF202300043

Certificate No.

DIRECTIONS

第 2 页, 共 3 页

Page of

1. 本中心是国家市场监督管理总局在华南地区设立的国家法定计量检定机构, 本中心的质量管理体系符合 ISO/IEC 17025:2017 标准的要求。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the General Administration of Quality Supervision. The quality system is in accordance with ISO/IEC 17025:2017.

2. 本中心所出具的数据均可溯源至国家计量基准和/或国际单位制(SI)。

All data issued by this laboratory are traceable to national primary standards and/or International System of Units (SI)

3. 校准地点、环境条件:

Place and environmental conditions of the calibration:

地点 本院热工恒温室 温度 (20.0~21.0) °C 相对湿度 60 %
Place Temperature RH

4. 本次校准的技术依据:

Reference documents for the calibration:

JJG 856-2015 工作用辐射温度计检定规程 V.R. of Radiation Thermometers

5. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号规格 Name of Equipment /Model/Type	编号 Serial No.	证书号/有效期/溯源单位 Certificate No./Due Date /Traceability to	计量特性 Metrological Characteristic
黑体辐射源/M340	B50510	RGW202202784 /2023-05-30 /省计量院	$U=(0.5\sim0.9)^{\circ}\text{C} (k=2)$
黑体辐射源/4181	B8B890	RGW202300412 /2024-01-08 /省计量院	$U=(0.5\sim2.1)^{\circ}\text{C} (k=2)$
红外测温仪/572-2	34561937WS	RGW202202789 /2023-05-30 /省计量院	MPE: $\pm 1.0\%$ 读数

注: 1. 本证书校准结果只与受校准仪器有关。 The results relate only to the items calibrated.

Note: 2. 未经本机构书面批准, 不得部分复制此证书。 This certificate shall not be reproduced except in full, without the written approval of our laboratory.

3. “委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“出厂编号”以及“设备编号”为仪器上标注。 The information Client and Contact Information are provided by client, and the Manufacturer, Model/Type, Serial No. and Equipment No. are marked on the items.

4. 本次校准日期视为发布日期。 The calibration date is the date of issue of the certificate.



校准结果 RESULTS OF CALIBRATION

证书编号 GRRF202300043
Certificate No.

原始记录号 202300043
Record No.

第 3 页, 共 3 页
Page of

- 一、外观检查: 符合要求。
二、光学系统: 符合要求。
三、固有误差: (单位: $^{\circ}\text{C}$)

校准点	参考温度计示值	被校温度计示值	固有误差	扩展不确定度 $U(k=2)$
-20	-20.0	-19.4	0.6	1.0
0	0.0	-0.2	-0.2	0.8
20	20.0	19.6	-0.4	0.6
50	49.9	49.8	-0.1	0.6
100	100.1	99.9	-0.2	1.0
150	150.1	149.2	-0.9	1.3
300	300.1	297.0	-3.1	1.8

说明:

- 被校辐射温度计的光谱范围: $(8\sim 14)\mu\text{m}$; 发射率设置为 $\varepsilon=0.95$ 。
温度范围 $(-20\sim \pm 20)^{\circ}\text{C}$, 校准距离: 5cm; 温度范围 $(50\sim 300)^{\circ}\text{C}$, 校准距离: 20cm。
- 本证书中给出的扩展不确定度依据 JJF1059.1-2012《测量不确定度评定与表示》评定, 由合成标准不确定度乘以包含概率约为 95% 时对应的包含因子 k 得到。
- 校准结果符合性判定依据 JJF1094-2002《测量仪器特性评定》之 5.3.1 和 JJG856-2015《工作用辐射温度计检定规程》。
- 按照所依据技术文件的规定, 建议复校时间间隔不超过 1 年。
更换重要部件、维修或对仪器性能有怀疑时, 应及时校准。

以下空白