

Test Verification of Conformity

Verification Number: 211230021GZU-VOC002

On the basis of the referenced test report(s), sample(s) tested of the below product have been found to comply with the standards harmonized with the regulation(s) listed on this verification at the time the tests were carried out. Other standards and Regulations may be relevant to the product. This verification is part of the full test report(s) and should be read in conjunction with it <them>.

Once compliance with all product relevant **UK CA** mark regulations are verified, including any relevant e.g. risk assessment and production control, the manufacturer may indicate compliance by signing a Declaration of Conformity themselves and applying the mark to products identical to the tested sample(s).

Applicant Name & Address:	Uni-Trend Technology (China) Co., Ltd No 6, Gong Ye Bei 1st Road, Songshan Lake National High-Tech Industrial Development Zone, Dongguan City, Guangdong Province, China
Product Description:	Infrared Thermometer
Models/Type References:	UT305C+, UT305A+, UTi32
Ratings & Principle Characteristics:	Type input: 5V, 1A; Powered by 1 x 3.7V 2600mAh 18650 rechargeable lithium battery
Brand Name:	UNI-T
Relevant Standards/Regulations:	EN 61326-1: 2013 EN 61326-2-3:2013 Electromagnetic Compatibility Regulations 2016
Verification Issuing Office Name & Address:	Intertek Testing Services Shenzhen Ltd. Guangzhou Branch Room 02, & 101/E201/E301/E401/E501/E601/E701/E801 of Room 01 1-8/F., No. 7-2. Caipin Road, Science City, GETDD, Guangzhou, Guangdong, China
Date of Tests:	30 December 2021 to 23 February 2022
Test Report Number(s):	211230021GZU-001

Sky Zhu

Signature

Name: Sky Zhu

Position: Team Leader

Date: 01 March 2022

This Verification is for the exclusive use of Intertek's client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Verification. Only the Client is authorized to permit copying or distribution of this Verification. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test/inspection results referenced in this Verification are relevant only to the sample tested/inspected. This Verification by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.